

CALIBRATION CERTIFICATE

Manufacturer : TERUMO

Certificate No. : -

Equipment : Blood Pressure Monitor

Hospital/Pharmacy : -

Model : BP500

Customer : -

Serial Number : BR21K02840

ID. No. : -

Environment : Ambient Temperature = -10 ~ 60 °C

Relative Humidity: 10 ~ 95%

Procedure used :

Reference Standard used:

ID.NO.	Type/Manufacturer	Model	S/N	Cert. No.	Due Date
	BP Pump-2M/Fluke	BP PUMP-2M	4453003	21MD178	2/10/2022

Inspection Result

referred to inspection report

General test

Visual inspection



Pass



Fail

Functional test



Pass



Fail

Electrical safety test



Pass



Fail

Ground wire : 0.08 Ω

Leakage Current : Nomal 124 μA

Reversed 224 μA

Performance test

Pressure measurement

Applied mmHg	Increasing		Applied mmHg	Decreasing	
	Reading	Error		Reading	Error
0	0	0	250	250	0
50	50	0	200	200	0
100	100	0	150	150	0
150	150	0	100	100	0
200	200	0	50	50	0
250	250	0	0	0	0

Air leakage rate

Reading (mmHg)	After (1 minute)	Result (mmHg/min)	Pass
250	249	1	☒

Over Pressure



N/A Over Pressure

Pressure (mmHg)	Reading	Pass	Fail
301	D06	☒	

※ D06 mean : Over Pressure

Calibration date(dd/mm/yyyy) : 23/11/2021

Due date(mm/yyyy) : -/11/2022

Perform by :

Orapan

(ORAPAN CHOOTONG)

QA Supervisor :

215

(NORASEJ PHONGSAI)



คู่มือการใช้งาน

BP500

เครื่องวัดความดันโลหิต เทอรูโม
รุ่น บีพี500 แบบอัตโนมัติ

